

# UK DECLARATION OF INCORPORATION

In accordance with UK Government guidance

This declaration is issued under the sole responsibility of the manufacturer,  
Heidolph Instruments GmbH & Co. KG  
Walpersdorfer Straße 12  
91126 Schwabach / Germany

Product: Laboratory automatic device

Model:

DISTIMATIC PRO HVI W. DRAINAGE A. ADAPT. A	591-6111x-xx-x
DISTIMATIC PRO HVI W. DRAINAGE A. ADAPT. R	591-6112x-xx-x
DISTIMATIC PRO HVI W/O DRAINAGE A. ADAPT. A	591-6121x-xx-x
DISTIMATIC PRO HVI W/O DRAINAGE A. ADAPT. R	591-6122x-xx-x

Description:

Device for filling and emptying of evaporation equipment.

The object of the declaration described above is classified as Partly Completed Machinery in accordance with UK Statutory Instrument 2008 No. 1597 Supply of Machinery (Safety) Regulations 2008.

The object of the declaration described above also is in conformity with the relevant UK Statutory Instruments (and their amendments):

2008 2016 No. 1091	The Electromagnetic Compatibility Regulations 2016
2012 No. 3032	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

This product is supplied as an optional accessory for incorporation to a Machine Tool. The device has to be installed in accordance with the instructions contained in the Installation Guide.

The partly completed machinery must not be put into service until it has been established that the machinery into which the partly completed machinery is to be incorporated complies with the provisions of a relevant product directive as a whole.

And complies with the following technical standards :

EN ISO 12100:2010, EN 61326-1:2013, EN 61010-1:2010, EN IEC 63000:2018

UK Authorised Representative (for authorities only):

ProductIP ( UK ) Ltd.  
8. Northumberland Av.  
London WC2N 5BY

Signed for and on behalf of Heidolph Instruments GmbH & Co. KG  
Walpersdorfer Straße 12, 91126 Schwabach / Germany

Schwabach, 17.08.2021



Wolfgang Jaenicke  
Managing Director



Jörg Ziel  
Quality Manager